

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 01235 Calib. Date: 12/27/2012 Calib. No.: 00022
Certification File No.: 01236 Cert. Date: 12/27/2012 Cert. No.: 00016
Linearity File No.: 01237 Lin. Date: 12/27/2012 Lin. No.: 00016
Solution File No.: 01247 Soln. Date: 01/21/2013 Soln. No.: 00153
Sequential File No.: 01247 File Date: 01/21/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 09/13/2014
Solution Control Lot: 121107 Bottle No.: 0927

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:05S		
Control 1 EC	0.100%	12:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:06S		
Control 2 EC	0.099%	12:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08S		
Control 3 EC	0.100%	12:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: Joseph Beni 543

Badge No.: 543

Date: 01/21/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01235 **Calib. Date:** 12/27/2012 **Calib. No.:** 00022
Certification File No.: 01236 **Cert. Date:** 12/27/2012 **Cert. No.:** 00016
Linearity File No.: 01237 **Lin. Date:** 12/27/2012 **Lin. No.:** 00016
Solution File No.: 01256 **Soln. Date:** 02/20/2013 **Soln. No.:** 00154
Sequential File No.: 01256 **File Date:** 02/20/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0928

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:58S		
Control 1 EC	0.101%	09:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		
Control 2 EC	0.099%	10:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01S		
Control 3 EC	0.099%	10:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: P+1 / 8m L -

543

Badge No.: 543

Date: 02/20/2013

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 01235 Calib. Date: 12/27/2012 Calib. No.: 00022
Certification File No.: 01236 Cert. Date: 12/27/2012 Cert. No.: 00016
Linearity File No.: 01237 Lin. Date: 12/27/2012 Lin. No.: 00016
Solution File No.: 01268 Soln. Date: 03/15/2013 Soln. No.: 00155
Sequential File No.: 01268 File Date: 03/15/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 09/13/2014
Solution Control Lot: 12I107 Bottle No.: 0929

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:18S		
Control 1 EC	0.101%	06:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:19S		
Control 2 EC	0.098%	06:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:20S		
Control 3 EC	0.098%	06:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	06:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature:  543

Badge No.: 543

Date: 03/15/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01235 **Calib. Date:** 12/27/2012 **Calib. No.:** 00022
Certification File No.: 01236 **Cert. Date:** 12/27/2012 **Cert. No.:** 00016
Linearity File No.: 01237 **Lin. Date:** 12/27/2012 **Lin. No.:** 00016
Solution File No.: 01278 **Soln. Date:** 04/05/2013 **Soln. No.:** 00156
Sequential File No.: 01278 **File Date:** 04/05/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0930

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:44S		
Control 1 EC	0.100%	14:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46S		
Control 2 EC	0.097%	14:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47S		
Control 3 EC	0.097%	14:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:49S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: 

543

Badge No.: 543

Date: 04/05/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01235 **Calib. Date:** 12/27/2012 **Calib. No.:** 00022
Certification File No.: 01236 **Cert. Date:** 12/27/2012 **Cert. No.:** 00016
Linearity File No.: 01237 **Lin. Date:** 12/27/2012 **Lin. No.:** 00016
Solution File No.: 01288 **Soln. Date:** 04/17/2013 **Soln. No.:** 00157
Sequential File No.: 01288 **File Date:** 04/17/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0932

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:30D		
Control 1 EC	0.101%	12:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31D		
Control 2 EC	0.099%	12:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32D		
Control 3 EC	0.099%	12:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: Joseph M Beni

Badge No.: 543

Date: 04/17/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01301 **Calib. Date:** 05/09/2013 **Calib. No.:** 00023
Certification File No.: 01302 **Cert. Date:** 05/09/2013 **Cert. No.:** 00017
Linearity File No.: 01303 **Lin. Date:** 05/09/2013 **Lin. No.:** 00017
Solution File No.: 01304 **Soln. Date:** 05/09/2013 **Soln. No.:** 00158
Sequential File No.: 01304 **File Date:** 05/09/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0933

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:09D		
Control 1 EC	0.101%	14:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:10D		
Control 2 EC	0.099%	14:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:11D		
Control 3 EC	0.100%	14:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJP2-144 (MA)

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: [Signature]

Badge No.: 6040

Date: 05/09/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	01301	Calib. Date:	05/09/2013
Certification File No.:	01302	Cert. Date:	05/09/2013
Linearity File No.:	01303	Lin. Date:	05/09/2013
Solution File No.:	01321	Soln. Date:	06/03/2013
Sequential File No.:	01321	File Date:	06/03/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	12I107	Expires:	09/13/2014
		Bottle No.:	0934

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:09D		
Control 1 EC	0.100%	10:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		
Control 2 EC	0.099%	10:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 3 EC	0.099%	10:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: P+1 Joseph 8/11/2013 543

Badge No.: 543

Date: 06/03/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01301 Calib. Date: 05/09/2013 Calib. No.: 00023
Certification File No.: 01302 Cert. Date: 05/09/2013 Cert. No.: 00017
Linearity File No.: 01303 Lin. Date: 05/09/2013 Lin. No.: 00017
Solution File No.: 01332 Soln. Date: 06/28/2013 Soln. No.: 00160
Sequential File No.: 01332 File Date: 06/28/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 09/13/2014
Solution Control Lot: 12I107 Bottle No.: 0935

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:37D		
Control 1 EC	0.100%	06:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:38D		
Control 2 EC	0.099%	06:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:39D		
Control 3 EC	0.099%	06:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: P+1 J. Dell Beni

543

Badge No.: 543

Date: 06/28/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01301 Calib. Date: 05/09/2013 Calib. No.: 00023
Certification File No.: 01302 Cert. Date: 05/09/2013 Cert. No.: 00017
Linearity File No.: 01303 Lin. Date: 05/09/2013 Lin. No.: 00017
Solution File No.: 01343 Soln. Date: 07/25/2013 Soln. No.: 00161
Sequential File No.: 01343 File Date: 07/25/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 09/13/2014
Solution Control Lot: 12I107 Bottle No.: 0936

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:01D		
Control 1 EC	0.099%	18:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:02D		
Control 2 EC	0.098%	18:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:03D		
Control 3 EC	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: _____

Badge No.: 543

Date: 07/25/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01301 **Calib. Date:** 05/09/2013 **Calib. No.:** 00023
Certification File No.: 01302 **Cert. Date:** 05/09/2013 **Cert. No.:** 00017
Linearity File No.: 01303 **Lin. Date:** 05/09/2013 **Lin. No.:** 00017
Solution File No.: 01357 **Soln. Date:** 08/18/2013 **Soln. No.:** 00162
Sequential File No.: 01357 **File Date:** 08/18/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0937

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:17D		
Control 1 EC	0.099%	10:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 2 EC	0.098%	10:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		
Control 3 EC	0.099%	10:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: _____



Badge No.: 543

Date: 08/18/2013

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 01301
Calib. Date: 05/09/2013
Calib. No.: 00023
Certification File No.: 01302
Cert. Date: 05/09/2013
Cert. No.: 00017
Linearity File No.: 01303
Lin. Date: 05/09/2013
Lin. No.: 00017
Solution File No.: 01372
Soln. Date: 09/17/2013
Soln. No.: 00163
Sequential File No.: 01372
File Date: 09/17/2013
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDUF S3-0065
Control Solution %: 0.100%
Expires: 09/13/2014
Solution Control Lot: 12I107
Bottle No.: 0938

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:18D		
Control 1 EC	0.098%	11:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19D		
Control 2 EC	0.098%	11:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		
Control 3 EC	0.098%	11:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JANKOWSKI

First Name: JOSEPH

MI: E

Signature:  #714

Badge No.: 714

Date: 09/17/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01301 **Calib. Date:** 05/09/2013 **Calib. No.:** 00023
Certification File No.: 01302 **Cert. Date:** 05/09/2013 **Cert. No.:** 00017
Linearity File No.: 01303 **Lin. Date:** 05/09/2013 **Lin. No.:** 00017
Solution File No.: 01383 **Soln. Date:** 10/11/2013 **Soln. No.:** 00164
Sequential File No.: 01383 **File Date:** 10/11/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0939

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:28D		
Control 1 EC	0.099%	11:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		
Control 2 EC	0.099%	11:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 3 EC	0.099%	11:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JANKOWSKI

First Name: JOSEPH

MI: E

Signature: 

Badge No.: 714

Date: 10/11/2013

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 01301 Calib. Date: 05/09/2013 Calib. No.: 00023
Certification File No.: 01302 Cert. Date: 05/09/2013 Cert. No.: 00017
Linearity File No.: 01303 Lin. Date: 05/09/2013 Lin. No.: 00017
Solution File No.: 01395 Soln. Date: 11/04/2013 Soln. No.: 00165
Sequential File No.: 01395 File Date: 11/04/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 09/13/2014
Solution Control Lot: 12I107 Bottle No.: 0941

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:34S		
Control 1 EC	0.099%	11:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35S		
Control 2 EC	0.097%	11:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:36S		
Control 3 EC	0.098%	11:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:38S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JANKOWSKI

First Name: JOSEPH

MI: E

Signature:  # 714

Badge No.: 714

Date: 11/04/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01396 **Calib. Date:** 11/05/2013 **Calib. No.:** 00024
Certification File No.: 01397 **Cert. Date:** 11/05/2013 **Cert. No.:** 00018
Linearity File No.: 01398 **Lin. Date:** 11/05/2013 **Lin. No.:** 00018
Solution File No.: 01399 **Soln. Date:** 11/05/2013 **Soln. No.:** 00166
Sequential File No.: 01399 **File Date:** 11/05/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0940

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:47S		
Control 1 EC	0.100%	15:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		
Control 2 EC	0.099%	15:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	15:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49S		
Control 3 EC	0.099%	15:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDSP2-0396 (JLS)

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: Sgt. Thomas J. Snyder / 45792

Badge No.: 5792

Date: 11/05/2013

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 01396 **Calib. Date:** 11/05/2013 **Calib. No.:** 00024
Certification File No.: 01397 **Cert. Date:** 11/05/2013 **Cert. No.:** 00018
Linearity File No.: 01398 **Lin. Date:** 11/05/2013 **Lin. No.:** 00018
Solution File No.: 01410 **Soln. Date:** 11/24/2013 **Soln. No.:** 00167
Sequential File No.: 01410 **File Date:** 11/24/2013

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 09/13/2014
Solution Control Lot: 12I107 **Bottle No.:** 0942

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:36S		
Control 1 EC	0.098%	21:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:38S		
Control 2 EC	0.097%	21:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:39S		
Control 3 EC	0.097%	21:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: _____

Badge No.: 543

Date: 11/24/2013

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 01396 Calib. Date: 11/05/2013 Calib. No.: 00024
Certification File No.: 01397 Cert. Date: 11/05/2013 Cert. No.: 00018
Linearity File No.: 01398 Lin. Date: 11/05/2013 Lin. No.: 00018
Solution File No.: 01419 Soln. Date: 12/14/2013 Soln. No.: 00168
Sequential File No.: 01419 File Date: 12/14/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 09/13/2014
Solution Control Lot: 121107 Bottle No.: 0943

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:32S		
Control 1 EC	0.097%	17:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:33S		
Control 2 EC	0.096%	17:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:35S		
Control 3 EC	0.097%	17:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:36S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: E

First Name: JOSEPH

MI: E

Signature: 

Badge No.: 714

Date: 12/14/2013