

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00067 Calib. Date: 06/29/2005 Calib. No.: 00003
Certification File No.: 00068 Cert. Date: 06/29/2005 Cert. No.: 00002
Linearity File No.: 00069 Lin. Date: 06/29/2005 Lin. No.: 00002
Solution File No.: 00152 Soln. Date: 01/05/2006 Soln. No.: 00021
Sequential File No.: 00152 File Date: 01/05/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 11/13/2006
Solution Control Lot: 04K008 Bottle No.: 0875

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:14S		
Control 1 EC	0.099%	17:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:15S		
Control 2 EC	0.095%	17:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17S		
Control 3 EC	0.096%	17:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: *John J. Bresnen*

Badge No.: 531

Date: 01/05/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00067 **Calib. Date:** 06/29/2005 **Calib. No.:** 00003
Certification File No.: 00068 **Cert. Date:** 06/29/2005 **Cert. No.:** 00002
Linearity File No.: 00069 **Lin. Date:** 06/29/2005 **Lin. No.:** 00002
Solution File No.: 00160 **Soln. Date:** 01/27/2006 **Soln. No.:** 00022
Sequential File No.: 00160 **File Date:** 01/27/2006

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 09/13/2007
Solution Control Lot: 05I020 **Bottle No.:** 0196

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:13S		
Control 1 EC	0.102%	21:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15S		
Control 2 EC	0.098%	21:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16S		
Control 3 EC	0.100%	21:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 01/27/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00067
Calib. Date: 06/29/2005
Calib. No.: 00003
Certification File No.: 00068
Cert. Date: 06/29/2005
Cert. No.: 00002
Linearity File No.: 00069
Lin. Date: 06/29/2005
Lin. No.: 00002
Solution File No.: 00170
Soln. Date: 02/15/2006
Soln. No.: 00023
Sequential File No.: 00170
File Date: 02/15/2006
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDUF S3-0064
Control Solution %: 0.100%
Expires: 09/13/2007
Solution Control Lot: 051020
Bottle No.: 0197

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:38S		
Control 1 EC	0.101%	11:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39S		
Control 2 EC	0.098%	11:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41S		
Control 3 EC	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 02/15/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00067
Calib. Date: 06/29/2005
Calib. No.: 00003
Certification File No.: 00068
Cert. Date: 06/29/2005
Cert. No.: 00002
Linearity File No.: 00069
Lin. Date: 06/29/2005
Lin. No.: 00002
Solution File No.: 00175
Soln. Date: 03/09/2006
Soln. No.: 00024
Sequential File No.: 00175
File Date: 03/09/2006

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 05L022
Model No.: CU-34
Serial No.: DDUF S3-0064
Expires: 12/15/2007
Bottle No.: 0079

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:59S		
Control 1 EC	0.102%	14:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		
Control 2 EC	0.098%	15:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02S		
Control 3 EC	0.099%	15:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 03/09/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00067 Calib. Date: 06/29/2005 Calib. No.: 00003
Certification File No.: 00068 Cert. Date: 06/29/2005 Cert. No.: 00002
Linearity File No.: 00069 Lin. Date: 06/29/2005 Lin. No.: 00002
Solution File No.: 00179 Soln. Date: 03/29/2006 Soln. No.: 00025
Sequential File No.: 00179 File Date: 03/29/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 12/15/2007
Solution Control Lot: 05J022 Bottle No.: 0080

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:15S		
Control 1 EC	0.101%	11:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16S		
Control 2 EC	0.098%	11:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		
Control 3 EC	0.098%	11:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 03/29/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	00067	Calib. Date:	06/29/2005
Certification File No.:	00068	Cert. Date:	06/29/2005
Linearity File No.:	00069	Lin. Date:	06/29/2005
Solution File No.:	00183	Soln. Date:	04/07/2006
Sequential File No.:	00183	File Date:	04/07/2006
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	05L022	Expires:	12/15/2007
		Bottle No.:	0084

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:04D		
Control 1 EC	0.101%	15:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05D		
Control 2 EC	0.097%	15:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 3 EC	0.098%	15:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature:



Badge No.: 531

Date: 04/07/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00067
Calib. Date: 06/29/2005
Calib. No.: 00003
Certification File No.: 00068
Cert. Date: 06/29/2005
Cert. No.: 00002
Linearity File No.: 00069
Lin. Date: 06/29/2005
Lin. No.: 00002
Solution File No.: 00189
Soln. Date: 04/27/2006
Soln. No.: 00027
Sequential File No.: 00189
File Date: 04/27/2006

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 05L022
Model No.: CU-34
Serial No.: DDUF S3-0064
Expires: 12/15/2007
Bottle No.: 0085

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:02D		
Control 1 EC	0.100%	11:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:04D		
Control 2 EC	0.098%	11:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:05D		
Control 3 EC	0.099%	11:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: John J. Bresnen

Badge No.: 531

Date: 04/27/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00067 Calib. Date: 06/29/2005 Calib. No.: 00003
Certification File No.: 00068 Cert. Date: 06/29/2005 Cert. No.: 00002
Linearity File No.: 00069 Lin. Date: 06/29/2005 Lin. No.: 00002
Solution File No.: 00194 Soln. Date: 05/15/2006 Soln. No.: 00028
Sequential File No.: 00194 File Date: 05/15/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 12/15/2007
Solution Control Lot: 05L022 Bottle No.: 0095

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:04D		
Control 1 EC	0.098%	16:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06D		
Control 2 EC	0.096%	16:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07D		
Control 3 EC	0.097%	16:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 05/15/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00067
Calib. Date: 06/29/2005
Calib. No.: 00003
Certification File No.: 00068
Cert. Date: 06/29/2005
Cert. No.: 00002
Linearity File No.: 00069
Lin. Date: 06/29/2005
Lin. No.: 00002
Solution File No.: 00207
Soln. Date: 06/09/2006
Soln. No.: 00031
Sequential File No.: 00207
File Date: 06/09/2006

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 05L022
Model No.: CU-34
Serial No.: DDUF S3-0064
Expires: 12/15/2007
Bottle No.: 0090

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:12D		
Control 1 EC	0.103%	17:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:13D		
Control 2 EC	0.101%	17:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14D		
Control 3 EC	0.101%	17:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature:



Badge No.: 531

Date: 06/09/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00215
Calib. Date: 06/27/2006
Calib. No.: 00004
Certification File No.: 00216
Cert. Date: 06/27/2006
Cert. No.: 00003
Linearity File No.: 00217
Lin. Date: 06/27/2006
Lin. No.: 00003
Solution File No.: 00229
Soln. Date: 07/17/2006
Soln. No.: 00034
Sequential File No.: 00229
File Date: 07/17/2006

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 05L022
Model No.: CU-34
Serial No.: DDUF S3-0065
Expires: 12/15/2007
Bottle No.: 0100

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:06D		
Control 1 EC	0.100%	17:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07D		
Control 2 EC	0.097%	17:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:08D		
Control 3 EC	0.097%	17:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: *John A. Bresnen*

Badge No.: 531

Date: 07/17/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00215
Calib. Date: 06/27/2006
Calib. No.: 00004
Certification File No.: 00216
Cert. Date: 06/27/2006
Cert. No.: 00003
Linearity File No.: 00217
Lin. Date: 06/27/2006
Lin. No.: 00003
Solution File No.: 00231
Soln. Date: 08/03/2006
Soln. No.: 00035
Sequential File No.: 00231
File Date: 08/03/2006

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 06C027
Model No.: CU-34
Serial No.: DDUF S3-0065
Expires: 03/03/2008
Bottle No.: 0827

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:33D		
Control 1 EC	0.100%	15:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34D		
Control 2 EC	0.098%	15:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35D		
Control 3 EC	0.098%	15:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 08/03/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00240 Soln. Date: 08/22/2006 Soln. No.: 00036
Sequential File No.: 00240 File Date: 08/22/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0828

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:45D		
Control 1 EC	0.100%	21:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:46D		
Control 2 EC	0.097%	21:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48D		
Control 3 EC	0.097%	21:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 08/22/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	00215	Calib. Date:	06/27/2006
Certification File No.:	00216	Cert. Date:	06/27/2006
Linearity File No.:	00217	Lin. Date:	06/27/2006
Solution File No.:	00246	Soln. Date:	09/08/2006
Sequential File No.:	00246	File Date:	09/08/2006
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	06C027	Expires:	03/03/2008
		Bottle No.:	0829

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:44D		
Control 1 EC	0.100%	14:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		
Control 2 EC	0.097%	14:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47D		
Control 3 EC	0.097%	14:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature:



Badge No.: 531

Date: 09/08/2006

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00215
Calib. Date: 06/27/2006
Calib. No.: 00004
Certification File No.: 00216
Cert. Date: 06/27/2006
Cert. No.: 00003
Linearity File No.: 00217
Lin. Date: 06/27/2006
Lin. No.: 00003
Solution File No.: 00247
Soln. Date: 09/26/2006
Soln. No.: 00038
Sequential File No.: 00247
File Date: 09/26/2006

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 06C027
Model No.: CU-34
Serial No.: DDUF S3-0065
Expires: 03/03/2008
Bottle No.: 0832

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:10D		
Control 1 EC	0.101%	18:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12D		
Control 2 EC	0.098%	18:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:13D		
Control 3 EC	0.098%	18:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 09/26/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00256 Soln. Date: 10/18/2006 Soln. No.: 00039
Sequential File No.: 00256 File Date: 10/18/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0833

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:09D		
Control 1 EC	0.101%	14:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:11D		
Control 2 EC	0.098%	14:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:12D		
Control 3 EC	0.098%	14:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 10/18/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00265 Soln. Date: 11/07/2006 Soln. No.: 00040
Sequential File No.: 00265 File Date: 11/07/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0834

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:09S		
Control 1 EC	0.101%	11:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10S		
Control 2 EC	0.097%	11:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 3 EC	0.097%	11:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 11/07/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00272 Soln. Date: 11/30/2006 Soln. No.: 00041
Sequential File No.: 00272 File Date: 11/30/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0837

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:14S		
Control 1 EC	0.099%	16:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15S		
Control 2 EC	0.098%	16:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S		
Control 3 EC	0.098%	16:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: *John J. Bresnen*

Badge No.: 531

Date: 11/30/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00281 Soln. Date: 12/18/2006 Soln. No.: 00043
Sequential File No.: 00281 File Date: 12/18/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0839

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:44S		
Control 1 EC	0.100%	16:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45S		
Control 2 EC	0.097%	16:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		
Control 3 EC	0.097%	16:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: *John J. Bresnen*

Badge No.: 531

Date: 12/18/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00283 Soln. Date: 12/22/2006 Soln. No.: 00044
Sequential File No.: 00283 File Date: 12/22/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0842

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:46S		
Control 1 EC	0.101%	15:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		
Control 2 EC	0.096%	15:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49S		
Control 3 EC	0.097%	15:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: John J. Bresnen

Badge No.: 531

Date: 12/22/2006