

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00287 Soln. Date: 01/10/2007 Soln. No.: 00045
Sequential File No.: 00287 File Date: 01/10/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0843

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:25S		
Control 1 EC	0.100%	18:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:27S		
Control 2 EC	0.096%	18:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:28S		
Control 3 EC	0.096%	18:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:30S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 01/10/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00215 Calib. Date: 06/27/2006 Calib. No.: 00004
Certification File No.: 00216 Cert. Date: 06/27/2006 Cert. No.: 00003
Linearity File No.: 00217 Lin. Date: 06/27/2006 Lin. No.: 00003
Solution File No.: 00291 Soln. Date: 01/13/2007 Soln. No.: 00046
Sequential File No.: 00291 File Date: 01/13/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0844

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:54S		
Control 1 EC	0.100%	14:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55S		
Control 2 EC	0.095%	14:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 3 EC	0.096%	14:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 01/13/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297 Calib. Date: 01/29/2007 Calib. No.: 00006
Certification File No.: 00298 Cert. Date: 01/29/2007 Cert. No.: 00004
Linearity File No.: 00299 Lin. Date: 01/29/2007 Lin. No.: 00004
Solution File No.: 00305 Soln. Date: 02/27/2007 Soln. No.: 00049
Sequential File No.: 00305 File Date: 02/27/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/03/2008
Solution Control Lot: 06C027 Bottle No.: 0848

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:43S		
Control 1 EC	0.099%	16:43S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44S		
Control 2 EC	0.099%	16:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45S		
Control 3 EC	0.099%	16:46S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 02/27/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297 **Calib. Date:** 01/29/2007 **Calib. No.:** 00006
Certification File No.: 00298 **Cert. Date:** 01/29/2007 **Cert. No.:** 00004
Linearity File No.: 00299 **Lin. Date:** 01/29/2007 **Lin. No.:** 00004
Solution File No.: 00316 **Soln. Date:** 03/28/2007 **Soln. No.:** 00050
Sequential File No.: 00316 **File Date:** 03/28/2007

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 03/03/2008
Solution Control Lot: 06C027 **Bottle No.:** 0849

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:15S		
Control 1 EC	0.098%	10:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		
Control 2 EC	0.097%	10:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		
Control 3 EC	0.099%	10:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature:



Badge No.: 531

Date: 03/28/2007

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00297
Calib. Date: 01/29/2007
Calib. No.: 00006
Certification File No.: 00298
Cert. Date: 01/29/2007
Cert. No.: 00004
Linearity File No.: 00299
Lin. Date: 01/29/2007
Lin. No.: 00004
Solution File No.: 00326
Soln. Date: 04/26/2007
Soln. No.: 00051
Sequential File No.: 00326
File Date: 04/26/2007

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 06K037

Model No.: CU-34

Serial No.: DDUF S3-0064
Expires: 11/17/2008
Bottle No.: 0701

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:26D		
Control 1 EC	0.099%	15:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27D		
Control 2 EC	0.098%	15:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29D		
Control 3 EC	0.098%	15:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 04/26/2007

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297
Certification File No.: 00298
Linearity File No.: 00299
Solution File No.: 00335
Sequential File No.: 00335

Serial No.: ARTL-0005

Calib. Date: 01/29/2007
Cert. Date: 01/29/2007
Lin. Date: 01/29/2007
Soln. Date: 05/23/2007
File Date: 05/23/2007

Calib. No.: 00006
Cert. No.: 00004
Lin. No.: 00004
Soln. No.: 00052

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 06K037

Model No.: CU-34

Serial No.: DDUF S3-0064
Expires: 11/17/2008
Bottle No.: 0702

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:14D		
Control 1 EC	0.098%	16:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15D		
Control 2 EC	0.098%	16:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16D		
Control 3 EC	0.099%	16:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: John F. Bresnen

Badge No.: 531

Date: 05/23/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297 Calib. Date: 01/29/2007 Calib. No.: 00006
Certification File No.: 00298 Cert. Date: 01/29/2007 Cert. No.: 00004
Linearity File No.: 00299 Lin. Date: 01/29/2007 Lin. No.: 00004
Solution File No.: 00346 Soln. Date: 06/21/2007 Soln. No.: 00053
Sequential File No.: 00346 File Date: 06/21/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 11/17/2008
Solution Control Lot: 06K037 Bottle No.: 0703

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:40D		
Control 1 EC	0.099%	11:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41D		
Control 2 EC	0.098%	11:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		
Control 3 EC	0.099%	11:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: *John J. Bresnen*

Badge No.: 531

Date: 06/21/2007

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	00297	Calib. Date:	01/29/2007
Certification File No.:	00298	Cert. Date:	01/29/2007
Linearity File No.:	00299	Lin. Date:	01/29/2007
Solution File No.:	00358	Soln. Date:	07/19/2007
Sequential File No.:	00358	File Date:	07/19/2007
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	06K037	Expires:	11/17/2008
		Bottle No.:	0704

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:55D		
Control 1 EC	0.099%	18:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:56D		
Control 2 EC	0.098%	18:56D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:57D		
Control 3 EC	0.098%	18:58D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature:



Badge No.: 531

Date: 07/19/2007

Calibrating Unit

New Standard Solution Report

Equipment

Location: Alcotest 7110 MKIII-C
PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297
Certification File No.: 00298
Linearity File No.: 00299
Solution File No.: 00362
Sequential File No.: 00362

Calib. Date: 01/29/2007
Cert. Date: 01/29/2007
Lin. Date: 01/29/2007
Soln. Date: 07/31/2007
File Date: 07/31/2007

Serial No.: ARTL-0005

Calib. No.: 00006
Cert. No.: 00004
Lin. No.: 00004
Soln. No.: 00055

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 06K037

Model No.: CU-34

Serial No.: DDUF S3-0065
Expires: 11/17/2008
Bottle No.: 0705

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:05D		
Control 1 EC	0.100%	12:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:06D		
Control 2 EC	0.099%	12:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07D		
Control 3 EC	0.099%	12:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUF S3-144 (592)

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: THOMAS J. SNYDER #5792

Badge No.: 5792

Date: 07/31/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297 Calib. Date: 01/29/2007 Calib. No.: 00006
Certification File No.: 00298 Cert. Date: 01/29/2007 Cert. No.: 00004
Linearity File No.: 00299 Lin. Date: 01/29/2007 Lin. No.: 00004
Solution File No.: 00379 Soln. Date: 08/28/2007 Soln. No.: 00056
Sequential File No.: 00379 File Date: 08/28/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 11/17/2008
Solution Control Lot: 06K037 Bottle No.: 0706

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:13D		
Control 1 EC	0.100%	12:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		
Control 2 EC	0.099%	12:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15D		
Control 3 EC	0.099%	12:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 08/28/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00297 Calib. Date: 01/29/2007 Calib. No.: 00006
Certification File No.: 00298 Cert. Date: 01/29/2007 Cert. No.: 00004
Linearity File No.: 00299 Lin. Date: 01/29/2007 Lin. No.: 00004
Solution File No.: 00390 Soln. Date: 09/26/2007 Soln. No.: 00057
Sequential File No.: 00390 File Date: 09/26/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 11/17/2008
Solution Control Lot: 06K037 Bottle No.: 0707

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:17D		
Control 1 EC	0.099%	12:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18D		
Control 2 EC	0.099%	12:19D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19D		
Control 3 EC	0.099%	12:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: 

Badge No.: 531

Date: 09/26/2007

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00297 Calib. Date: 01/29/2007 Calib. No.: 00006
Certification File No.: 00298 Cert. Date: 01/29/2007 Cert. No.: 00004
Linearity File No.: 00299 Lin. Date: 01/29/2007 Lin. No.: 00004
Solution File No.: 00399 Soln. Date: 10/26/2007 Soln. No.: 00058
Sequential File No.: 00399 File Date: 10/26/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 11/17/2008
Solution Control Lot: 06K037 Bottle No.: 0708

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:58D		
Control 1 EC	0.100%	11:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 2 EC	0.098%	12:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		
Control 3 EC	0.099%	12:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature:



Badge No.: 531

Date: 10/26/2007

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 00403
Calib. Date: 11/14/2007
Calib. No.: 00007
Certification File No.: 00404
Cert. Date: 11/14/2007
Cert. No.: 00005
Linearity File No.: 00405
Lin. Date: 11/14/2007
Lin. No.: 00005
Solution File No.: 00406
Soln. Date: 11/14/2007
Soln. No.: 00059
Sequential File No.: 00406
File Date: 11/14/2007
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDUF S3-0065
Control Solution %: 0.100%
Expires: 11/17/2008
Solution Control Lot: 06K037
Bottle No.: 0709

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:18S		
Control 1 EC	0.101%	12:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19S		
Control 2 EC	0.099%	12:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21S		
Control 3 EC	0.099%	12:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUSP2-144 (5792)

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: TPR. II Thomas J. Snyder #5792

Badge No.: 5792

Date: 11/14/2007

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 00403 Calib. Date: 11/14/2007 Calib. No.: 00007
Certification File No.: 00404 Cert. Date: 11/14/2007 Cert. No.: 00005
Linearity File No.: 00405 Lin. Date: 11/14/2007 Lin. No.: 00005
Solution File No.: 00417 Soln. Date: 12/13/2007 Soln. No.: 00060
Sequential File No.: 00417 File Date: 12/13/2007

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 11/17/2008
Solution Control Lot: 06K037 Bottle No.: 0710

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:25S		
Control 1 EC	0.099%	18:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:26S		
Control 2 EC	0.097%	18:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:28S		
Control 3 EC	0.098%	18:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRESNEN

First Name: JOHN

MI: A

Signature: *John J. Bresnen*

Badge No.: 531

Date: 12/13/2007