

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02119 **Calib. Date:** 12/09/2020 **Calib. No.:** 00046
Certification File No.: 02120 **Cert. Date:** 12/09/2020 **Cert. No.:** 00037
Linearity File No.: 02121 **Lin. Date:** 12/09/2020 **Lin. No.:** 00037
Solution File No.: 02127 **Soln. Date:** 01/19/2021 **Soln. No.:** 00285
Sequential File No.: 02127 **File Date:** 01/19/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 02/19/2022
Solution Control Lot: 20090 **Bottle No.:** 0374

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:10S		
Control 1 EC	0.100%	11:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 2 EC	0.099%	11:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 3 EC	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: Richard R. Wolak #554

Badge No.: 554

Date: 01/19/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02119 **Calib. Date:** 12/09/2020 **Calib. No.:** 00046
Certification File No.: 02120 **Cert. Date:** 12/09/2020 **Cert. No.:** 00037
Linearity File No.: 02121 **Lin. Date:** 12/09/2020 **Lin. No.:** 00037
Solution File No.: 02131 **Soln. Date:** 03/19/2021 **Soln. No.:** 00287
Sequential File No.: 02131 **File Date:** 03/19/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 01/13/2023
Solution Control Lot: 21010 **Bottle No.:** 0025

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:58S		
Control 1 EC	0.102%	10:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59S		
Control 2 EC	0.101%	11:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	11:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01S		
Control 3 EC	0.101%	11:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____



Badge No.: 554

Date: 03/19/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02119	Calib. Date:	12/09/2020
Certification File No.:	02120	Cert. Date:	12/09/2020
Linearity File No.:	02121	Lin. Date:	12/09/2020
Solution File No.:	02139	Soln. Date:	04/17/2021
Sequential File No.:	02139	File Date:	04/17/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	0085

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:18D		
Control 1 EC	0.100%	15:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 2 EC	0.100%	15:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		
Control 3 EC	0.100%	15:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	15:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak

Badge No.: 554

Date: 04/17/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02119 Calib. Date: 12/09/2020 Calib. No.: 00046
Certification File No.: 02120 Cert. Date: 12/09/2020 Cert. No.: 00037
Linearity File No.: 02121 Lin. Date: 12/09/2020 Lin. No.: 00037
Solution File No.: 02148 Soln. Date: 05/29/2021 Soln. No.: 00289
Sequential File No.: 02148 File Date: 05/29/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0077

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:59D		
Control 1 EC	0.102%	14:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00D		
Control 2 EC	0.101%	14:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 3 EC	0.102%	14:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak

Badge No.: 554

Date: 05/29/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02163 Calib. Date: 07/30/2021 Calib. No.: 00050
Certification File No.: 02164 Cert. Date: 07/30/2021 Cert. No.: 00038
Linearity File No.: 02165 Lin. Date: 07/30/2021 Lin. No.: 00038
Solution File No.: 02166 Soln. Date: 07/30/2021 Soln. No.: 00293
Sequential File No.: 02166 File Date: 07/30/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0084

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:25D		
Control 1 EC	0.102%	11:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 2 EC	0.101%	11:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 3 EC	0.101%	11:28D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number:

DDUJ P2 - 144

MRW

Changed By:

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature:

Tpr. I  # 2078

Badge No.: 7078

Date: 07/30/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02163 Calib. Date: 07/30/2021 Calib. No.: 00050
Certification File No.: 02164 Cert. Date: 07/30/2021 Cert. No.: 00038
Linearity File No.: 02165 Lin. Date: 07/30/2021 Lin. No.: 00038
Solution File No.: 02172 Soln. Date: 08/28/2021 Soln. No.: 00294
Sequential File No.: 02172 File Date: 08/28/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0032

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:02D		
Control 1 EC	0.099%	18:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	18:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:03D		
Control 2 EC	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:05D		
Control 3 EC	0.099%	18:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____



Badge No.: 554

Date: 08/28/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02163 **Calib. Date:** 07/30/2021 **Calib. No.:** 00050
Certification File No.: 02164 **Cert. Date:** 07/30/2021 **Cert. No.:** 00038
Linearity File No.: 02165 **Lin. Date:** 07/30/2021 **Lin. No.:** 00038
Solution File No.: 02181 **Soln. Date:** 10/12/2021 **Soln. No.:** 00295
Sequential File No.: 02181 **File Date:** 10/12/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0065
Control Solution %: 0.100% **Expires:** 03/17/2023
Solution Control Lot: 21090 **Bottle No.:** 1103

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:22D		
Control 1 EC	0.099%	16:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23D		
Control 2 EC	0.098%	16:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:24D		
Control 3 EC	0.098%	16:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: 

Badge No.: 554

Date: 10/12/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02163 Calib. Date: 07/30/2021 Calib. No.: 00050
Certification File No.: 02164 Cert. Date: 07/30/2021 Cert. No.: 00038
Linearity File No.: 02165 Lin. Date: 07/30/2021 Lin. No.: 00038
Solution File No.: 02186 Soln. Date: 11/16/2021 Soln. No.: 00296
Sequential File No.: 02186 File Date: 11/16/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1139

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:37S		
Control 1 EC	0.100%	08:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38S		
Control 2 EC	0.098%	08:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40S		
Control 3 EC	0.098%	08:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak #554

Badge No.: 554

Date: 11/16/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02163 Calib. Date: 07/30/2021 Calib. No.: 00050
Certification File No.: 02164 Cert. Date: 07/30/2021 Cert. No.: 00038
Linearity File No.: 02165 Lin. Date: 07/30/2021 Lin. No.: 00038
Solution File No.: 02189 Soln. Date: 12/20/2021 Soln. No.: 00297
Sequential File No.: 02189 File Date: 12/20/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0021

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:43S		
Control 1 EC	0.101%	12:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:45S		
Control 2 EC	0.100%	12:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:46S		
Control 3 EC	0.098%	12:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OBRIEN

First Name: DARBY

MI: F

Signature: 

Badge No.: 566

Date: 12/20/2021

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 02190
Calib. Date: 12/21/2021
Calib. No.: 00051
Certification File No.: 02191
Cert. Date: 12/21/2021
Cert. No.: 00039
Linearity File No.: 02192
Lin. Date: 12/21/2021
Lin. No.: 00039
Solution File No.: 02193
Soln. Date: 12/21/2021
Soln. No.: 00298
Sequential File No.: 02193
File Date: 12/21/2021

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 21090
Model No.: CU-34
Serial No.: DDUF S3-0064
Expires: 03/17/2023
Bottle No.: 1154

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:55S		
Control 1 EC	0.101%	14:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 2 EC	0.100%	14:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58S		
Control 3 EC	0.098%	14:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number:

DDSF P2-0396 (CG)

Changed By:

Last Name: GUARCELLO

First Name: COREY

MI: R.

Signature:

[Signature] #7035

Badge No.: 7035

Date: 12/21/2021