

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02196 Soln. Date: 01/17/2022 Soln. No.: 00299
Sequential File No.: 02196 File Date: 01/17/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1149

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:29S		
Control 1 EC	-.---%	08:29S	23.0°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	08:29S	23.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: Richard Wolak

Badge No.: 554

Date: 01/17/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02197 Soln. Date: 01/19/2022 Soln. No.: 00300
Sequential File No.: 02197 File Date: 01/19/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1455

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:55S		
Control 1 EC	0.099%	14:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56S		
Control 2 EC	0.098%	14:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 3 EC	0.098%	14:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

Badge No.: 554

Date: 01/19/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02200 Soln. Date: 02/13/2022 Soln. No.: 00301
Sequential File No.: 02200 File Date: 02/13/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1493

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:03S		
Control 1 EC	0.100%	14:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04S		
Control 2 EC	0.099%	14:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:06S		
Control 3 EC	0.100%	14:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

Richard Wolak

Badge No.: 554

Date: 02/13/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02203 Soln. Date: 03/13/2022 Soln. No.: 00302
Sequential File No.: 02203 File Date: 03/13/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0640

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:21S		
Control 1 EC	0.098%	15:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22S		
Control 2 EC	0.097%	15:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24S		
Control 3 EC	0.097%	15:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R WOLAK

Badge No.: 554

Date: 03/13/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 **Calib. Date:** 12/21/2021 **Calib. No.:** 00051
Certification File No.: 02191 **Cert. Date:** 12/21/2021 **Cert. No.:** 00039
Linearity File No.: 02192 **Lin. Date:** 12/21/2021 **Lin. No.:** 00039
Solution File No.: 02207 **Soln. Date:** 04/11/2022 **Soln. No.:** 00303
Sequential File No.: 02207 **File Date:** 04/11/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0064
Control Solution %: 0.100% **Expires:** 07/21/2023
Solution Control Lot: 21270 **Bottle No.:** 0692

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:43D		
Control 1 EC	0.099%	14:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		
Control 2 EC	0.097%	14:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		
Control 3 EC	0.097%	14:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak #554

Badge No.: 554

Date: 04/11/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02209 Soln. Date: 05/10/2022 Soln. No.: 00304
Sequential File No.: 02209 File Date: 05/10/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0603

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:59D		
Control 1 EC	0.099%	13:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01D		
Control 2 EC	0.096%	13:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	13:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 3 EC	0.096%	13:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R A R Wolak

Badge No.: 554

Date: 05/10/2022

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PLAINSBORO TOWNSHIP PD
Serial No.: ARTL-0005
Calibration File No.: 02212 Calib. Date: 05/24/2022 Calib. No.: 00052
Certification File No.: 02213 Cert. Date: 05/24/2022 Cert. No.: 00040
Linearity File No.: 02214 Lin. Date: 05/24/2022 Lin. No.: 00040
Solution File No.: 02215 Soln. Date: 05/24/2022 Soln. No.: 00305
Sequential File No.: 02215 File Date: 05/24/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0398

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:08D		
Control 1 EC	0.102%	13:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09D		
Control 2 EC	0.101%	13:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10D		
Control 3 EC	0.100%	13:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-144 (CS)

Changed By:

Last Name: GUARCELLO

First Name: COREY

MI: R.

Signature:  7035

Badge No.: 7035

Date: 05/24/2022

Calibrating Unit

New Standard Solution Report

Equipment

Location: Alcotest 7110 MKIII-C
PLAINSBORO TOWNSHIP PD

Serial No.: ARTL-0005

Calibration File No.: 02212

Calib. Date: 05/24/2022

Calib. No.: 00052

Certification File No.: 02213

Cert. Date: 05/24/2022

Cert. No.: 00040

Linearity File No.: 02214

Lin. Date: 05/24/2022

Lin. No.: 00040

Solution File No.: 02219

Soln. Date: 06/24/2022

Soln. No.: 00306

Sequential File No.: 02219

File Date: 06/24/2022

Calibrating Unit: WET

Model No.: CU-34

Serial No.: DDUF S3-0065

Control Solution %: 0.100%

Expires: 07/21/2023

Solution Control Lot: 21270

Bottle No.: 1400

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:22D		
Control 1 EC	0.099%	11:23D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 2 EC	0.098%	11:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		
Control 3 EC	0.097%	11:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature:

RA R Wolak #554

Badge No.: 554

Date: 06/24/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02212	Calib. Date:	05/24/2022
Certification File No.:	02213	Calib. No.:	00052
Linearity File No.:	02214	Cert. Date:	05/24/2022
Solution File No.:	02224	Cert. No.:	00040
Sequential File No.:	02224	Lin. Date:	05/24/2022
		Lin. No.:	00040
		Soln. Date:	07/21/2022
		Soln. No.:	00307
		File Date:	07/21/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	21270	Expires:	07/21/2023
		Bottle No.:	0602

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:46D		
Control 1 EC	0.100%	13:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47D		
Control 2 EC	0.098%	13:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		
Control 3 EC	0.098%	13:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak

Badge No.: 554

Date: 07/21/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02212	Calib. Date:	05/24/2022
Certification File No.:	02213	Cert. Date:	05/24/2022
Linearity File No.:	02214	Lin. Date:	05/24/2022
Solution File No.:	02227	Soln. Date:	08/04/2022
Sequential File No.:	02227	File Date:	08/04/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	22070	Expires:	02/17/2024
		Bottle No.:	0760

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:09D		
Control 1 EC	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		
Control 2 EC	0.099%	10:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 3 EC	0.099%	10:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: 

Badge No.: 743

Date: 08/04/2022