

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD	
Calibration File No.:	02003	Calib. Date: 12/26/2018
Certification File No.:	02004	Calib. No.: 00039
Linearity File No.:	02005	Cert. Date: 12/26/2018
Solution File No.:	02010	Cert. No.: 00031
Sequential File No.:	02010	Lin. Date: 12/26/2018
		Lin. No.: 00031
		Soln. Date: 01/15/2019
		Soln. No.: 00260
		File Date: 01/15/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUF S3-0064
Control Solution %:	0.100%		Expires: 05/10/2020
Solution Control Lot:	18150		Bottle No.: 0779

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:24S		
Control 1 EC	0.099%	12:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26S		
Control 2 EC	0.098%	12:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27S		
Control 3 EC	0.098%	12:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: 

Badge No.: 543

Date: 01/15/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02003	Calib. Date:	12/26/2018
Certification File No.:	02004	Cert. Date:	12/26/2018
Linearity File No.:	02005	Lin. Date:	12/26/2018
Solution File No.:	02017	Soln. Date:	02/08/2019
Sequential File No.:	02017	File Date:	02/08/2019
Calibrating Unit: WET		Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	0303

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:45S		
Control 1 EC	0.099%	16:46S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		
Control 2 EC	0.097%	16:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48S		
Control 3 EC	0.098%	16:49S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_

cpl J. Beni

543

Badge No.: 543

Date: 02/08/2019

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02003 Calib. Date: 12/26/2018 Calib. No.: 00039  
Certification File No.: 02004 Cert. Date: 12/26/2018 Cert. No.: 00031  
Linearity File No.: 02005 Lin. Date: 12/26/2018 Lin. No.: 00031  
Solution File No.: 02022 Soln. Date: 03/06/2019 Soln. No.: 00262  
Sequential File No.: 02022 File Date: 03/06/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064  
Control Solution %: 0.100% Expires: 05/10/2020  
Solution Control Lot: 18150 Bottle No.: 0706

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:32S		
Control 1 EC	0.099%	12:33S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34S		
Control 2 EC	0.096%	12:34S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35S		
Control 3 EC	0.097%	12:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	12:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_

*R. R. Wolak*

Badge No.: 554

Date: 03/06/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02003	Calib. Date:	12/26/2018
Certification File No.:	02004	Cert. Date:	12/26/2018
Linearity File No.:	02005	Lin. Date:	12/26/2018
Solution File No.:	02028	Soln. Date:	04/11/2019
Sequential File No.:	02028	File Date:	04/11/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	0540

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:11D		
Control 1 EC	0.097%	14:11D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:12D		
Control 2 EC	0.095%	14:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	14:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14D		
Control 3 EC	0.096%	14:14D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_

CPL J. Dell Beni

543

Badge No.: 543

Date: 04/11/2019



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02003	Calib. Date:	12/26/2018
Certification File No.:	02004	Cert. Date:	12/26/2018
Linearity File No.:	02005	Lin. Date:	12/26/2018
Solution File No.:	02035	Soln. Date:	05/26/2019
Sequential File No.:	02035	File Date:	05/26/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	1038

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:26D		
Control 1 EC	0.097%	20:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27D		
Control 2 EC	0.097%	20:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:29D		
Control 3 EC	0.097%	20:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_

*[Handwritten Signature]* 543

Badge No.: 543

Date: 05/26/2019

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02003	Calib. Date:	12/26/2018
Certification File No.:	02004	Calib. No.:	00039
Linearity File No.:	02005	Cert. Date:	12/26/2018
Solution File No.:	02040	Cert. No.:	00031
Sequential File No.:	02040	Lin. Date:	12/26/2018
		Lin. No.:	00031
		Soln. Date:	06/10/2019
		Soln. No.:	00267
		File Date:	06/10/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	1035

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:28D		
Control 1 EC	0.097%	13:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29D		
Control 2 EC	0.096%	13:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:30D		
Control 3 EC	0.097%	13:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_



Badge No.: 543

Date: 06/10/2019

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02041 Calib. Date: 06/12/2019 Calib. No.: 00040  
Certification File No.: 02042 Cert. Date: 06/12/2019 Cert. No.: 00032  
Linearity File No.: 02043 Lin. Date: 06/12/2019 Lin. No.: 00032  
Solution File No.: 02044 Soln. Date: 06/12/2019 Soln. No.: 00268  
Sequential File No.: 02044 File Date: 06/12/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065  
Control Solution %: 0.100% Expires: 02/11/2021  
Solution Control Lot: 19060 Bottle No.: 1059

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:12D		
Control 1 EC	0.101%	10:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		
Control 2 EC	0.100%	10:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		
Control 3 EC	0.100%	10:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-144 

**Changed By:**

Last Name: MIMIKOS

First Name: NICHOLAS

MI: E.

Signature: TPR.  #7413

Badge No.: 7413

Date: 06/12/2019



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02041	Calib. Date:	06/12/2019
Certification File No.:	02042	Cert. Date:	06/12/2019
Linearity File No.:	02043	Lin. Date:	06/12/2019
Solution File No.:	02052	Soln. Date:	07/20/2019
Sequential File No.:	02052	File Date:	07/20/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	1058

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:42D		
Control 1 EC	0.100%	21:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:44D		
Control 2 EC	0.099%	21:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:45D		
Control 3 EC	0.099%	21:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: \_\_\_\_\_

*col J Beni - 543*

Badge No.: 543

Date: 07/20/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD .	
Calibration File No.:	02041	Calib. Date: 06/12/2019
Certification File No.:	02042	Cert. No.: 00040
Linearity File No.:	02043	Cert. Date: 06/12/2019
Solution File No.:	02057	Lin. No.: 00032
Sequential File No.:	02057	Soln. No.: 00270
		File Date: 08/18/2019
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0065
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 1055

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:50D		
Control 1 EC	0.099%	15:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51D		
Control 2 EC	0.098%	15:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:53D		
Control 3 EC	0.098%	15:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_

R. Wolak #554

Badge No.: 554

Date: 08/18/2019

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02041	Calib. Date:	06/12/2019
Certification File No.:	02042	Cert. No.:	00040
Linearity File No.:	02043	Cert. Date:	06/12/2019
Solution File No.:	02062	Lin. Date:	06/12/2019
Sequential File No.:	02062	Lin. No.:	00032
		Soln. Date:	09/13/2019
		Soln. No.:	00271
		File Date:	09/13/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0065
Control Solution %:	0.100%			Expires:	02/11/2021
Solution Control Lot:	19060			Bottle No.:	1036

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:48D		
Control 1 EC	0.099%	15:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49D		
Control 2 EC	0.098%	15:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50D		
Control 3 EC	0.098%	15:51D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_



Badge No.: 554

Date: 09/13/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02041	Calib. Date:	06/12/2019
Certification File No.:	02042	Cert. No.:	00040
Linearity File No.:	02043	Cert. Date:	06/12/2019
Solution File No.:	02069	Lin. No.:	00032
Sequential File No.:	02069	Lin. Date:	06/12/2019
		Soln. No.:	00032
		Soln. Date:	10/19/2019
		File Date:	10/19/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	1137

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:09D		
Control 1 EC	0.099%	19:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:10D		
Control 2 EC	0.098%	19:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:12D		
Control 3 EC	0.098%	19:12D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_

*R.H. R. Wolak #554*

Badge No.: 554

Date: 10/19/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02041	Calib. Date:	06/12/2019
Certification File No.:	02042	Cert. Date:	06/12/2019
Linearity File No.:	02043	Lin. Date:	06/12/2019
Solution File No.:	02074	Soln. Date:	11/19/2019
Sequential File No.:	02074	File Date:	11/19/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0065
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0812

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:22S		
Control 1 EC	0.098%	07:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 2 EC	0.097%	07:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25S		
Control 3 EC	0.097%	07:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_

*Richard W. Wolak*

Badge No.: 554

Date: 11/19/2019