

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARTL-0005 |
| Location: | PLAINSBORO TOWNSHIP PD | | |
| Calibration File No.: | 02083 | Calib. Date: | 01/22/2020 |
| Certification File No.: | 02084 | Cert. Date: | 01/22/2020 |
| Linearity File No.: | 02085 | Lin. Date: | 01/22/2020 |
| Solution File No.: | 02093 | Soln. Date: | 02/16/2020 |
| Sequential File No.: | 02093 | File Date: | 02/16/2020 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUF S3-0064 |
| Solution Control Lot: | 18290 | Expires: | 09/17/2020 |
| | | Bottle No.: | 1257 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 14:22S | | |
| Control 1 EC | 0.099% | 14:22S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 14:22S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:23S | | |
| Control 2 EC | 0.098% | 14:24S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 14:24S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:24S | | |
| Control 3 EC | 0.098% | 14:25S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 14:25S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:26S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: Richard Wolak

Badge No.: 554

Date: 02/16/2020

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02083 Calib. Date: 01/22/2020 Calib. No.: 00044
Certification File No.: 02084 Cert. Date: 01/22/2020 Cert. No.: 00035
Linearity File No.: 02085 Lin. Date: 01/22/2020 Lin. No.: 00035
Solution File No.: 02100 Soln. Date: 04/14/2020 Soln. No.: 00277
Sequential File No.: 02100 File Date: 04/14/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0825

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 08:52D | | |
| Control 1 EC | 0.099% | 08:53D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 08:53D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:54D | | |
| Control 2 EC | 0.098% | 08:54D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 08:54D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:55D | | |
| Control 3 EC | 0.099% | 08:56D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 08:56D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:56D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: Richard Wolak #554

Badge No.: 554

Date: 04/14/2020

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0005 |
| Location: | PLAINSBORO TOWNSHIP PD | |
| Calibration File No.: | 02083 | Calib. Date: 01/22/2020 |
| Certification File No.: | 02084 | Cert. No.: 00044 |
| Linearity File No.: | 02085 | Cert. Date: 01/22/2020 |
| Solution File No.: | 02101 | Lin. No.: 00035 |
| Sequential File No.: | 02101 | Soln. Date: 06/03/2020 |
| | | Soln. No.: 00278 |
| | | File Date: 06/03/2020 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUF S3-0064 |
| Solution Control Lot: | 18290 | Expires: 09/17/2020 |
| | | Bottle No.: 1012 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 09:35D | | |
| Control 1 EC | 0.100% | 09:35D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 09:35D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:36D | | |
| Control 2 EC | 0.099% | 09:36D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 09:36D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:37D | | |
| Control 3 EC | 0.099% | 09:38D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 09:38D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:38D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R WOLAK

Badge No.: 554

Date: 06/03/2020

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0005 |
| Location: | PLAINSBORO TOWNSHIP PD | |
| Calibration File No.: | 02102 | Calib. Date: 07/01/2020 |
| Certification File No.: | 02103 | Calib. No.: 00045 |
| Linearity File No.: | 02104 | Cert. Date: 07/01/2020 |
| Solution File No.: | 02105 | Lin. No.: 00036 |
| Sequential File No.: | 02105 | Soln. No.: 00279 |
| | | File Date: 07/01/2020 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUF S3-0065 |
| Solution Control Lot: | 20090 | Expires: 02/19/2022 |
| | | Bottle No.: 0430 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 13:22D | | |
| Control 1 EC | 0.102% | 13:22D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 13:22D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:23D | | |
| Control 2 EC | 0.101% | 13:24D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 13:24D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:24D | | |
| Control 3 EC | 0.101% | 13:25D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 13:25D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:26D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-144 MPW

Changed By:

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature:  #2078

Badge No.: 7078

Date: 07/01/2020

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARTL-0005 |
| Location: | PLAINSBORO TOWNSHIP PD | | |
| Calibration File No.: | 02102 | Calib. Date: | 07/01/2020 |
| Certification File No.: | 02103 | Cert. Date: | 07/01/2020 |
| Linearity File No.: | 02104 | Lin. Date: | 07/01/2020 |
| Solution File No.: | 02109 | Soln. Date: | 08/03/2020 |
| Sequential File No.: | 02109 | File Date: | 08/03/2020 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUF S3-0065 |
| Solution Control Lot: | 18290 | Expires: | 09/17/2020 |
| | | Bottle No.: | 0892 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 20:40D | | |
| Control 1 EC | 0.099% | 20:41D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 20:41D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:42D | | |
| Control 2 EC | 0.098% | 20:42D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 20:42D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:43D | | |
| Control 3 EC | 0.099% | 20:43D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 20:43D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:44D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELL BENI

First Name: JOSEPH

MI: M

Signature: _____

[Handwritten Signature] 743

Badge No.: 743

Date: 08/03/2020

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02102 Calib. Date: 07/01/2020 Calib. No.: 00045
Certification File No.: 02103 Cert. Date: 07/01/2020 Cert. No.: 00036
Linearity File No.: 02104 Lin. Date: 07/01/2020 Lin. No.: 00036
Solution File No.: 02110 Soln. Date: 08/10/2020 Soln. No.: 00281
Sequential File No.: 02110 File Date: 08/10/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 1136

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:48D | | |
| Control 1 EC | 0.099% | 07:49D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 07:49D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:49D | | |
| Control 2 EC | 0.098% | 07:50D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 07:50D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:51D | | |
| Control 3 EC | 0.099% | 07:51D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 07:51D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:52D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: 

Badge No.: 554

Date: 08/10/2020

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARTL-0005 |
| Location: | PLAINSBORO TOWNSHIP PD | | |
| Calibration File No.: | 02102 | Calib. Date: | 07/01/2020 |
| Certification File No.: | 02103 | Cert. Date: | 07/01/2020 |
| Linearity File No.: | 02104 | Lin. Date: | 07/01/2020 |
| Solution File No.: | 02112 | Soln. Date: | 09/21/2020 |
| Sequential File No.: | 02112 | File Date: | 09/21/2020 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUF S3-0065 |
| Solution Control Lot: | 20090 | Expires: | 02/19/2022 |
| | | Bottle No.: | 0147 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 09:28D | | |
| Control 1 EC | 0.100% | 09:29D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 09:29D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:30D | | |
| Control 2 EC | 0.099% | 09:30D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 09:30D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:31D | | |
| Control 3 EC | 0.099% | 09:32D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 09:32D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:32D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak #554

Badge No.: 554

Date: 09/21/2020

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02102 Calib. Date: 07/01/2020 Calib. No.: 00045
Certification File No.: 02103 Cert. Date: 07/01/2020 Cert. No.: 00036
Linearity File No.: 02104 Lin. Date: 07/01/2020 Lin. No.: 00036
Solution File No.: 02115 Soln. Date: 10/30/2020 Soln. No.: 00283
Sequential File No.: 02115 File Date: 10/30/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065
Control Solution %: 0.100% Expires: 02/19/2022
Solution Control Lot: 20090 Bottle No.: 0373

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 20:50S | | |
| Control 1 EC | 0.100% | 20:51S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 20:51S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:52S | | |
| Control 2 EC | 0.099% | 20:52S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 20:52S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:53S | | |
| Control 3 EC | 0.099% | 20:54S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 20:54S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:54S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

PTL R. Wolak #554

Badge No.: 554

Date: 10/30/2020

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARTL-0005 |
| Location: | PLAINSBORO TOWNSHIP PD | | |
| Calibration File No.: | 02119 | Calib. Date: | 12/09/2020 |
| Certification File No.: | 02120 | Cert. Date: | 12/09/2020 |
| Linearity File No.: | 02121 | Lin. Date: | 12/09/2020 |
| Solution File No.: | 02122 | Soln. Date: | 12/09/2020 |
| Sequential File No.: | 02122 | File Date: | 12/09/2020 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUF S3-0064 |
| Solution Control Lot: | 20090 | Expires: | 02/19/2022 |
| | | Bottle No.: | 0667 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:24S | | |
| Control 1 EC | 0.102% | 10:24S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 10:24S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:25S | | |
| Control 2 EC | 0.101% | 10:26S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 10:26S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:27S | | |
| Control 3 EC | 0.100% | 10:27S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 10:27S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:28S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDSP2-0396 MRW

Changed By:

Last Name: WATSON First Name: MATTHEW MI: R

Signature:  #2078 Badge No.: 7078
 Date: 12/09/2020