

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD	
Calibration File No.:	02119	Calib. No.: 00046
Certification File No.:	02120	Cert. No.: 00037
Linearity File No.:	02121	Lin. No.: 00037
Solution File No.:	02127	Soln. No.: 00285
Sequential File No.:	02127	File Date: 01/19/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0064
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0374

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:10S		
Control 1 EC	0.100%	11:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 2 EC	0.099%	11:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 3 EC	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature:           *Rtl R. Wolak #554*          

Badge No.: 554

Date: 01/19/2021

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02119 Calib. Date: 12/09/2020 Calib. No.: 00046  
Certification File No.: 02120 Cert. Date: 12/09/2020 Cert. No.: 00037  
Linearity File No.: 02121 Lin. Date: 12/09/2020 Lin. No.: 00037  
Solution File No.: 02131 Soln. Date: 03/19/2021 Soln. No.: 00287  
Sequential File No.: 02131 File Date: 03/19/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0025

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:58S		
Control 1 EC	0.102%	10:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59S		
Control 2 EC	0.101%	11:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	11:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01S		
Control 3 EC	0.101%	11:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_



Badge No.: 554

Date: 03/19/2021

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02119 Calib. Date: 12/09/2020 Calib. No.: 00046  
Certification File No.: 02120 Cert. Date: 12/09/2020 Cert. No.: 00037  
Linearity File No.: 02121 Lin. Date: 12/09/2020 Lin. No.: 00037  
Solution File No.: 02139 Soln. Date: 04/17/2021 Soln. No.: 00288  
Sequential File No.: 02139 File Date: 04/17/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0085

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:18D		
Control 1 EC	0.100%	15:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 2 EC	0.100%	15:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		
Control 3 EC	0.100%	15:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	15:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_



Badge No.: 554

Date: 04/17/2021

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02119 Calib. Date: 12/09/2020 Calib. No.: 00046  
Certification File No.: 02120 Cert. Date: 12/09/2020 Cert. No.: 00037  
Linearity File No.: 02121 Lin. Date: 12/09/2020 Lin. No.: 00037  
Solution File No.: 02148 Soln. Date: 05/29/2021 Soln. No.: 00289  
Sequential File No.: 02148 File Date: 05/29/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0077

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:59D		
Control 1 EC	0.102%	14:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00D		
Control 2 EC	0.101%	14:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 3 EC	0.102%	14:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_

*R. R. Wolak*

Badge No.: 554

Date: 05/29/2021

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD	
Calibration File No.:	02163	Calib. Date: 07/30/2021
Certification File No.:	02164	Calib. No.: 00050
Linearity File No.:	02165	Cert. Date: 07/30/2021
Solution File No.:	02166	Cert. No.: 00038
Sequential File No.:	02166	Lin. Date: 07/30/2021
		Lin. No.: 00038
		Soln. Date: 07/30/2021
		Soln. No.: 00293
		File Date: 07/30/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0065
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0084

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:25D		
Control 1 EC	0.102%	11:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 2 EC	0.101%	11:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 3 EC	0.101%	11:28D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2 - 144 MRW

**Changed By:**

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature:  # 2078

Badge No.: 7078

Date: 07/30/2021

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02163 Calib. Date: 07/30/2021 Calib. No.: 00050  
Certification File No.: 02164 Cert. Date: 07/30/2021 Cert. No.: 00038  
Linearity File No.: 02165 Lin. Date: 07/30/2021 Lin. No.: 00038  
Solution File No.: 02172 Soln. Date: 08/28/2021 Soln. No.: 00294  
Sequential File No.: 02172 File Date: 08/28/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0032

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:02D		
Control 1 EC	0.099%	18:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	18:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:03D		
Control 2 EC	0.099%	18:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:05D		
Control 3 EC	0.099%	18:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_



Badge No.: 554

Date: 08/28/2021

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0005  
Location: PLAINSBORO TOWNSHIP PD  
Calibration File No.: 02163 Calib. Date: 07/30/2021 Calib. No.: 00050  
Certification File No.: 02164 Cert. Date: 07/30/2021 Cert. No.: 00038  
Linearity File No.: 02165 Lin. Date: 07/30/2021 Lin. No.: 00038  
Solution File No.: 02181 Soln. Date: 10/12/2021 Soln. No.: 00295  
Sequential File No.: 02181 File Date: 10/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0065  
Control Solution %: 0.100% Expires: 03/17/2023  
Solution Control Lot: 21090 Bottle No.: 1103

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:22D		
Control 1 EC	0.099%	16:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23D		
Control 2 EC	0.098%	16:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:24D		
Control 3 EC	0.098%	16:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: \_\_\_\_\_



Badge No.: 554

Date: 10/12/2021