

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD	
Calibration File No.:	02190	Calib. Date: 12/21/2021
Certification File No.:	02191	Calib. No.: 00051
Linearity File No.:	02192	Cert. Date: 12/21/2021
Solution File No.:	02196	Cert. No.: 00039
Sequential File No.:	02196	Lin. Date: 12/21/2021
		Lin. No.: 00039
		Soln. Date: 01/17/2022
		Soln. No.: 00299
		File Date: 01/17/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0064
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 1149

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:29S		
Control 1 EC	-.---%	08:29S	23.0°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	08:29S	23.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: 

Badge No.: 554

Date: 01/17/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD	
Calibration File No.:	02190	Calib. Date: 12/21/2021
Certification File No.:	02191	Calib. No.: 00051
Linearity File No.:	02192	Cert. Date: 12/21/2021
Solution File No.:	02197	Cert. No.: 00039
Sequential File No.:	02197	Lin. Date: 12/21/2021
		Lin. No.: 00039
		Soln. Date: 01/19/2022
		Soln. No.: 00300
		File Date: 01/19/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0064
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 1455

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:55S		
Control 1 EC	0.099%	14:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56S		
Control 2 EC	0.098%	14:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 3 EC	0.098%	14:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____



Badge No.: 554

Date: 01/19/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02200 Soln. Date: 02/13/2022 Soln. No.: 00301
Sequential File No.: 02200 File Date: 02/13/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 1493

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:03S		
Control 1 EC	0.100%	14:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04S		
Control 2 EC	0.099%	14:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:06S		
Control 3 EC	0.100%	14:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

Richard Wolak

Badge No.: 554

Date: 02/13/2022

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0005
Location: PLAINSBORO TOWNSHIP PD
Calibration File No.: 02190 Calib. Date: 12/21/2021 Calib. No.: 00051
Certification File No.: 02191 Cert. Date: 12/21/2021 Cert. No.: 00039
Linearity File No.: 02192 Lin. Date: 12/21/2021 Lin. No.: 00039
Solution File No.: 02203 Soln. Date: 03/13/2022 Soln. No.: 00302
Sequential File No.: 02203 File Date: 03/13/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0064
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0640

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:21S		
Control 1 EC	0.098%	15:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22S		
Control 2 EC	0.097%	15:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24S		
Control 3 EC	0.097%	15:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

RWR Wolak

Badge No.: 554

Date: 03/13/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD		
Calibration File No.:	02190	Calib. Date:	12/21/2021
Certification File No.:	02191	Cert. No.:	00051
Linearity File No.:	02192	Cert. Date:	12/21/2021
Solution File No.:	02207	Lin. No.:	00039
Sequential File No.:	02207	Lin. Date:	12/21/2021
		Soln. Date:	04/11/2022
		Soln. No.:	00303
		File Date:	04/11/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0064
Solution Control Lot:	21270	Expires:	07/21/2023
		Bottle No.:	0692

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:43D		
Control 1 EC	0.099%	14:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		
Control 2 EC	0.097%	14:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		
Control 3 EC	0.097%	14:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R. Wolak #554

Badge No.: 554

Date: 04/11/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0005
Location:	PLAINSBORO TOWNSHIP PD	
Calibration File No.:	02190	Calib. Date: 12/21/2021
Certification File No.:	02191	Calib. No.: 00051
Linearity File No.:	02192	Cert. Date: 12/21/2021
Solution File No.:	02209	Cert. No.: 00039
Sequential File No.:	02209	Lin. Date: 12/21/2021
		Lin. No.: 00039
		Soln. Date: 05/10/2022
		Soln. No.: 00304
		File Date: 05/10/2022

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUF S3-0064
Control Solution %: 0.100%		Expires: 07/21/2023
Solution Control Lot: 21270		Bottle No.: 0603

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:59D		
Control 1 EC	0.099%	13:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01D		
Control 2 EC	0.096%	13:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	13:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 3 EC	0.096%	13:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLAK

First Name: RICHARD

MI: S

Signature: _____

R A R Wolak

Badge No.: 554

Date: 05/10/2022

